




## 5 Suggested guidelines for therapy

### Standard treatment procedure

#### Adapting to the patient

For a patient using this device for the first time, it is advisable to begin with lower pressures, such as 10 – 15 cmH<sub>2</sub>O positive and negative pressure, and low inhale flow. It will familiarize the patient with the feel of mechanical insufflation-exsufflation.



As the patient becomes more comfortable with the therapy, progressively increase the inspiratory and expiratory pressures by 5 to 10 cmH<sub>2</sub>O each sequence of 4 to 6 breaths. Effective pressures may be around 35 – 45 cmH<sub>2</sub>O.

**Tip:** To help the patient understand and become comfortable with the therapy, the patient synchronization dial is provided in the center of the screen during Auto mode.

**Cough-Trak** may also be enabled to help synchronize the therapy with the patient's inspiratory effort to make acclimation to the therapy easier.


#### General therapy description


- 1 cough cycle is composed of an inspiratory, expiratory and pause phase
- 4 to 6 cough cycles compose a sequence
- 20 to 30 seconds of resting time is suggested between sequences allowing for secretion removal
- A treatment is composed of 4 to 6 sequences and can be performed several times a day, whenever needed
- Preferably, therapy should be performed in the morning, before meals and bedtime

	Feature	Adults	Pediatrics
Basics	Effective pressure (cmH <sub>2</sub> O)	+/- 35 up to +/- 45	+/- 45 adjust to patient condition
	Inhale time (in s)	2 to 3	1 to 2
	Exhale time (in s)	2 to 3	2 or less
	Pause time (in s) In Auto mode	2 to 3	As small as possible
Customize	Cough-Trak triggering feature In Auto mode	For patients who have some inspiratory drive this feature will trigger on the patient's inspiration to help synchronize the therapy with the patient's effort.	
	Mode and Preset	Therapy parameters can be titrated/modified in Manual mode to evaluate the best therapy for the patient before setting up automatic settings  Up to 3 different Presets can be saved to provide different therapies (i.e. prophylactic use, mucus plugs, lung volume recruitment, etc.)	
Follow-up	Data management	Data such as <b>peak cough flow, tidal volume, SpO<sub>2</sub></b> , or time counters, available in detailed view <b>ON, can be useful for both setting therapy parameters and assessing therapy efficiency</b>	
	DirectView EncorePro	Therapy data recorded (internal memory and SD card): peak cough flow, tidal volume, pressure, SpO <sub>2</sub> , heart rate, pattern of usage	

**!** The settings should always be checked before beginning therapy. Guidelines are intended to serve only as reference. They do not supersede established medical protocols or instructions from physician.

## 6 Device menu

To access device menu, press the up button  below the menu. The menu selection and their corresponding parameters are displayed in the table below:

Options	LCD brightness, date format, time format, month, day, year, hour and minutes. Only in Full access mode: menu access, detailed view, language, pressure units, therapy hours
Data	SpO <sub>2</sub> , heart rate (HR), peak cough flow, tidal volume, serial number, model number, software version, calibration date, SD card capacity, detach battery serial number, detach battery cycles, therapy hours
Information Log	Informational messages about the device will be indicated by the  icon. The log menu displays the list of messages that occurred, i.e. 'battery not charging' or 'replace battery' and allows them to be cleared
Clear Patient Data	Allows clearing previous data <b>from device memory and SD card</b> to record new data (i.e. for a new patient)
Safely Remove SD Card	Allows removing SD card without risk of losing or damaging recorded information; follow messages displayed on screen
Write Event Log to SD Card	Copies event log data from the device to the SD Card; follow message displayed on screen



## 7 Mask fitting

To choose a well fitted mask, ensure the entire mouth and nose are included inside the mask.


Make sure the top of the cushion is not too high on the bridge of the nose area. No leak should remain.

#### Landmarks

- a Side of mouth
- b Bridge of nose zone
- c Below the lower lip



Mask size (inches)	Length	Width	Height
Infant	2.76	2.60	1.93
Toddler	3.46	3.15	2.32
Small adult	4.53	3.74	3.19
Medium adult	4.84	4.02	3.54
Large adult	4.92	4.33	3.54



## CoughAssist T70





## Quick start guide

- 1 Device overview
- 2 Starting therapy
- 3 Modifying therapy settings
- 4 Monitoring view and icons
- 5 Suggested guidelines for therapy
- 6 Device menu
- 7 Mask fitting

# 1 Device overview

- 1 Display screen
- 2 Manual switch to activate inhale (right) and exhale (left) phases
- 3 Power on/Power off
- 4 Perform action specified on screen
- 5 Navigate menus or change setup




- 6 Breathing circuit connection
- 7 Fan exhaust – Air outlet must be kept uncovered
- 8 Detachable battery
- 9 Filter – Air inlet
- 10 AC power inlet
- 11 Remote control connector
- 12 USB connector (for service only)
- 13 SpO<sub>2</sub> connector
- 14 DC power inlet

**Note:** When the remote control foot pedal is connected, the manual switch is disabled.


# 2 Starting therapy

## How to start/stop therapy



### To start therapy

- Connect the device to AC or DC power source or use the detachable battery
- Press to power on the device
- Select **Therapy** to start the treatment




### To stop therapy

- Select **Standby**
- To turn power off, press

## How to lock/unlock


Limited access (lock) permits the user to select between Preset therapy settings, while Full access (unlock) will allow the user to change all settings.



### To temporarily unlock the device

- While in Standby screen, press the down button below **Menu** and hold manual switch to the right for 5 seconds

**Note:** When appears on the main screen, the device is unlocked. If the Access menu is not set to Full, the device will switch automatically to limited access after 5 minutes of inactivity.




### To select access option

- Press the up button below **Menu**
- While 'Option' is highlighted in green press **Select**
- While 'Menu Access' is highlighted, press **Modify** and then **Edit** to select 'Limited' or 'Full' access
- Press **OK** to validate
- Press **Finish** and **Exit** to return to the main screen

# 3 Modifying therapy settings

## To modify settings

- The therapy settings are accessible through the left button below **Settings**
- The scroll buttons allow selection of desired parameter
- Pressing **Modify** will allow changing the parameter highlighted in green, **OK** validates the choice and **Finish** returns to main screen
- Three Presets are available to save a group of settings for different circumstances



### The settings available are displayed in the table below:

Preset	1, 2, 3
Mode	Auto, Manual
Cough-Trak	OFF/ON <small>Only available in Auto mode</small>
Inhale pressure	0 to 70 cmH <sub>2</sub> O, in increments of 1 cmH <sub>2</sub> O
Inhale flow	Low, Medium, High
Inhale time	0 to 5s, in increments of 0.1s <small>Only available in Auto mode</small>
Exhale pressure	0 to -70 cmH <sub>2</sub> O, in increments of 1 cmH <sub>2</sub> O
Exhale time	0 to 5s, in increments of 0.1s <small>Only available in Auto mode</small>
Pause time	0 to 5s, in increments of 0.1s <small>Only available in Auto mode with Cough-Trak OFF</small>


**Note:** Any change of settings made during therapy will automatically take effect at the beginning of that phase in the next cycle.

# 4 Monitoring view and icons

## To change monitoring view

- Press the up button below **Menu**
- While 'Option' is highlighted in green, press **Select** and navigate down to 'Detailed View'
- Press **Modify** and **Edit** to change 'OFF' or 'ON'
- Press **OK** to validate
- Press **Finish** and **Exit** to return to the main screen


### Detailed View 'ON'



**Note:** When selecting detailed view 'OFF', the following parameters are not displayed:

- Peak Cough Flow
- SpO<sub>2</sub>
- HR
- Tidal Volume
- Exhale and Inhale Time setting and counter
- Pause Time setting

### Detailed View 'OFF'



## Symbols and icons

- Color code:
  - green = pause
  - yellow = exhale
  - blue = inhale
- Pulse oximeter: green means connected (flashing)
- With ? means questionable data
- SD card inserted
- SD card error
- Memory card writing: copying data to SD card
- Full access
- Informational message: information available in the Information Log
- Remote control: (foot pedal) connected
- External battery indicator: red = few minutes left
- Power source in use indicator: box is placed around the battery being used
- Detachable battery indicator: red = few minutes left
- Charging indicator: detachable battery is being charged